

CONTINUOUS INFUSION PUMP INSTRUCTIONS

The Infusion pump is working properly when:

- There is a green light flashing in the circle in the center of the pump
- There is a periodic audible movement of the gears (winding sound)

The pump is **NOT** working properly when:

- It alarms (could be short beeps or a constant tone)
- "OCL" indicates there is an occlusion (constant beeping):
 - To resolve, check the tubing for any kinks
 - Push the run/pause button to silence the alarm
 - Push the run/pause button again to restart the pump
 - If pump continues to alarm, call the chemo clinic
- "MA" indicates there is a malfunction with the pump (constant tone):
 - To resolve:
 - Push the run/pause button to silence the alarm
 - Gently press on the top of the cassette (top of pump)
 - Push the run/pause button again to restart the pump
- "Low battery" will be indicated by a blinking battery symbol on the screen
 - To resolve:
 - Push the run/pause button to silence the alarm
 - Replace the batteries immediately (2 AA batteries, found in pump bag)
 - To replace batteries, turn the bottom of pump all the way to the left until the cap comes off
 - After replacing batteries, replace bottom cap and turn to the right until the two lines pair up
 - Push the run/pause button again to restart the pump

If all attempts to correct the pump problem are exhausted, or if you need assistance, please call **1-800-444-2728** for further instructions.

Clinic hours are Monday – Friday, 8:30 am to 5:00 pm.

If it is after hours, or a weekend, and you have not been able to resolve the pump issue, please call (253) 428-8700 to page the Doctor on call for further instructions.

(See the next page for a diagram of the Infusion Pump)

AMBIT CONTINUOUS INFUSION PUMP DIAGRAM

